

Centro Nacional de Metrología

SIM Regional Supplementary Comparison

SIM.L-S6

Calibration of Gauge Blocks by Mechanical Comparison

Final Report

October 2014

- M. Viliesid
 CENAM, Centro Nacional de Metrología
- C. Colín
 CENAM, Centro Nacional de Metrología
- T. Chávez
 CENAM, Centro Nacional de Metrología
- K.P. Chaudhary
 NPLI, National Physical Laboratory INDIA
- F. Dvořáček
 CMI, Czech Metrology Institute
- J. Stoup NIST, National Institute of Standards and Technology
- W. Santos INMETRO, Instituto Nacional de Metrologia
- L. Vaudagna INTI, Instituto Nacional de Tecnología Industrial
- R. Morales
 DICTUC, Laboratorio Nacional de Longitud
- A. Acquarone
 LATU, Laboratorio Tecnológico del Uruguay
- J. Carrasco INDECOPI, Servicio Nacional de Metrología

- M. Vega IBMETRO, Instituto Boliviano de Metrología
- M. Salazar INEN, Instituto Ecuatoriano de Normalización
- V. Gil SIC, Superintendencia de Industria y Comercio
- J. Dimas
 CENAMEP, Centro Nacional de Metrología de Panamá AIP
- E. Reyes
 LACOMET, Laboratorio Costarricense de Metrología
- F. Hamilton
 TTBS, Trinidad and Tobago Bureau of
 Standards
- T. Reddock
 TTBS, Trinidad and Tobago Bureau of
 Standards
- S. Durga BSJ, Bureau of Standards Jamaica
- T. Burton
 BSJ, Bureau of Standards Jamaica

Centro Nacional de Metrología (CENAM)

km 4.5 Carretera a Los Cués, El Marqués, Querétaro 76246, MEXICO

1. Introduction

The Mutual Recognition Arrangement (MRA) of the *Conférence Internationale des Poids et Mesures* (*CIPM*) signed by the National Metrology Institutes (NMI) of different nations provides mutual recognition among the NMI of their national standards and their calibration services. A database has been set up by the *Bureau Interantional des Poids et Mesures* (*BIPM*) at its website where the Calibration and Measurement Capabilities (CMC) of each NMI are posted. To support the CMC claims of the NMI, the MRA requires, among other things, that they participate on a regular basis in Key Comparisons (KC) that test key measuring techniques. This would prove their technical competence, that they can provide this calibration service with the claimed uncertainty of the corresponding CMC and that they have metrological equivalence with the other signatory NMI that provide the same service.

The CIPM has therefore instructed the different CC (*Comité Consultatif*) to identify key techniques in order to define KC, as it is, for example, the calibration of Gauge Blocks (GB) by optical interferometry, identified as a key measuring technique by CCL (CC de *Longueurs*). Additionally, the CC as well as the regions may also identify other important comparisons called supplementary and identified with an S. The SIM region has identified the Calibration of GB by Mechanical Comparison as one if this comparisons. The *Centro Nacional de Metrología* (CENAM) was designated by SIM as pilot laboratory and CENAM has carried out this exercise under the name of SIM.L-S6:2007.

The calibration of GB by Mechanical Comparison is indeed a technique of paramount importance as it is at the highest level in the traceability chain of length for most countries of the American Continent. This comparison is meant to support the submitted CMC of these countries.

The measurand is the central length of the GB as defined in [1] and the circulated GB were used for two comparisons carried out in two stages:

- First stage, SIM.L- K1:2007, Calibration of GB by Optical Interferometry. Circulation from 2007-11-01 to 2010-04-25. The GB were also measured by Mechanical Comparison for those NMI also participating in SIM.L-S6:2007.
- Second stage, SIM.L-S6:2007, Calibration of GB by Mechanical Comparison. Circulation from 2010-03-02 to 2011-05-06 for the participants that measured only by Mechanical Comparison, this dates included two control measurements by interferometry.

In this second comparison there were 16 participants, 14 from the Americas, and 2 invited NMI from other regions. The circulation in the second stage had 10 participants. It should be noted that the circulation took relatively short time for the large number of participants thanks to the hand delivery from one NMI to the following one in nine out of the 10 steps so that customs clearance was simplified or avoided.

2. Participants

This comparison had 16 participants. **Table 1** shows the participating NMI and their corresponding contact person and information.

| Contact | NMI | Information |
|---------------------------------------|--|---|
| Carlos Colín | CENAM, Centro Nacional de Metrología | Tel. +52 442 211 0500 |
| Castellanos | km 4.5 Carretera a los Cués, El Marqués | Fax +52 442 211 0577 |
| | CP 76241, Querétaro, MÉXICO. | e-mail: ccolin@cenam.mx |
| K. P. Chaudhary | NPLI, National Physical Laboratory INDIA | Tel. +91 11 25732865 |
| | Dr. K,S. Krishnan Road, New Delhi | Fax +91 11 25726938 |
| | 110012, INDIA . | e-mail: kpc@mail.nplindia.ernet.in |
| Ing. Vladimir | CMI, Czech Metrology Institute | Tel. +42 485 107 532 |
| Stezka | Slunecna 23 | Fax +42 485 104 466 |
| Ing. Frantisek | 460 01 Liberec | e-mail: vstezka@cmi.cz |
| Dvořáček | CZECH REPUBLIC. | fdvoracek@cmi.cz |
| Leandro Vaudagna | INTI, Instituto Nacional de Tecnología Industrial | Tel. +54 34 92440471 |
| | División de Metroplogía Dimensional-Rafaela | Fax +54 34 92422804 |
| | km 227,6 Ruta Nac. N° 34 | e-mail: vaudagna@inti.gov.ar |
| | CP 2300, Rafaela, Santa Fé, ARGENTINA | |
| Wellington Santos | INMETRO, Instituto Nacional de Metrologia, | Tel. +55 21 2679-9271 |
| Barros | Normalização e Qualidade Industrial. | Fax +55 21 2679-9207 |
| | Av. N.Sra. das Graças, 50 – Villa Operária – | e-mail: wsbarros@inmetro.go.br |
| | Xerém – Duque de Caixas – RJ. CEP 25250- | |
| | 020, BRASIL. | |
| John Stoup | NIST, National Institute of Standards and | Tel. +1 301 975 3476 |
| | Technology | Fax + 1 301 869 0822 |
| | Room B113, Metrology Building | e-mail: John.Stoup@nist.gov |
| D 1 1 M 1 | Gaithersburg, MD 20899-0001 USA | T 50 0 0544004 |
| Roberto Morales | DICTUC, Laboratorio Nacional de Longitud | Tel. 56 2 3544624 |
| | Avenida Vicuña Mackenna 4860 – Macul – | Fax 56 2 3544624 |
| Alaiandra | Santiago – (edificio nº 9 metrología), CHILE. LATU, Laboratorio Tecnológico del Uruguay | e-mail: metrologia@dictuc.cl |
| Alejandro Acquarone | Avenida Italia 6201, Montevideo, URUGUAY . | Tel. 598 2 601 3724 ext 298 Fax 598 2 601 8554 |
| Luis Mussio | CP 11500 | e-mail: lmussio@latu.org.uy |
| Luis Mussio | CF 11300 | aacqua@latu.org.uy |
| Janet Carrasco | INDECOPI, Servicio Nacional de Metrología | Tel. 51 1 224 7800 1618 |
| Tuesta | Calle de la Prosa 138, San Borja | Fax 51 1 224 7800 1264 |
| | Lima 4,1 PERÚ . | e-mail: jcarrasco@indecopi.gob.pe |
| María del Carmen | IBMETRO, Instituto Boliviano de Metrología | Tel. 591 2 2372046 |
| Vega Amonzabel | Av. Camacho No. 1488 | Fax 591 2 2310037 |
| | La Paz, BOLIVIA . | e-mail: mvega@ibmetro.gob.bo |
| Manuel Salazar | INEN, Instituto Ecuatoriano de Normalización | Tel. 593 2 2343716 |
| | Casilla: 17-01-3999 | Fax 593 2 2344394 |
| \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ | Quito – ECUADOR. | e-mail: msalazar@inen.gov.ec |
| Victor Hugo Gil | SIC, Superintendencia de Industria y Comercio | Tel. 571 3153265 - 69 |
| | División de Metrología | Fax 571 3153292 |
| Julio Dimos | COLOMBIA. CENAMEP, Centro Nacional de Metrología de | e-mail: vgil@correo.sic.co Tel. 507 – 517-0081 |
| Julio Dimas | Panamá AIP, Edificio 215, Ciudad del Saber, | Fax 507 – 517-0081 |
| | PANAMÁ. | e-mail: jdimas@cenamep.org.pa |
| Eduardo Reyes | LACOMET, Laboratorio Costarricense de | Tel. 506 - 2283 - 6580 |
| | Metrología. | Fax 506 - 2283 - 5133 |
| | Apartado Postal 1736 – 2050, San Pedro de | e-mail: ereyes@lacomet.go.cr |
| | Montes de Oca, San Jose. COSTA RICA. | |
| Theodore Reddock | TTBS, Trinidad and Tobago Bureau of | Tel. 868-662-8827 |
| Francis Hamilton | Standards | Fax 868-663-4335 |
| | Century Drive Trincity Industrial Estate Macoya, | e-mail: Theodore.Reddock@ttbs.org.tt |
| | Tunapuna, TRINIDAD AND TOBAGO. | e-mail:Francis.Hamilton@ttbs.org.tt |
| Tomokie Burton | BSJ, Bureau of Standards, Jamaica | Tel.: 926-3140-5 ext. 1102 |
| Carlton Thomas | 6 Winchester Rd., | Fax: 929-4736 |
| Siew Durga | Kingston 10. JAMAICA . | e-mail: cthomas@bsj.org.jm |
| | Table 1. List of participants in comparison 9 | e-mail: TBurton@bsj.org.jm |

Tabla 1. List of participants in comparison SIM.L-S6:2007.

3. Circulation Schedule

As we mentioned, nine out of the ten participants in the second circulation stage delivered the artifacts by hand to the following participant which made the circulation time relatively short. Table 2 shows the circulation schedule for the two circulation stages. *LACOMET* retained 13 weeks the GB because of end-of-the-year holidays and internal administrative problems, they mentioned.

| NMI | Da | ates | Reception of results | | | | | |
|---------------|-----------------------|-----------------|----------------------|--|--|--|--|--|
| INIVII | Reception | Shipment | Reception of results | | | | | |
| | Stage C | One Circulation | | | | | | |
| NPLI | 2008-03-02 | 2008-06-09 | 2009-08-07 | | | | | |
| CMI | 2008-06-16 | 2008-08-05 | 2008-10-17 | | | | | |
| NIST | 2009-04-20 | 2009-07-15 | 2010-03-02 | | | | | |
| INMETRO | 2009-08-11 | 2009-09-08 | 2009-09-21 | | | | | |
| INTI | 2009-11-13 | 2010-01-11 | 2010-01-11 | | | | | |
| | Stage Two Circulation | | | | | | | |
| DICTUC | 2010-04-29 | 2010-05-18 | 2010-06-15 | | | | | |
| LATU | 2010-06-11 | 2010-07-06 | 2010-07-21 | | | | | |
| INDECOPI | 2010-07-06 | 2010-07-30 | 2010-08-12 | | | | | |
| IBMETRO | 2010-08-02 | 2010-08-18 | 2010-09-24 | | | | | |
| INEN | 2010-08-18 | 2010-09-13 | 2011-05-16 | | | | | |
| SIC | 2010-09-13 | 2010-10-13 | 2010-11-10 | | | | | |
| CENAMEP | 2010-10-13 | 2010-11-09 | 2011-01-28 | | | | | |
| LACOMET | 2010-11-09 | 2011-02-09 | 2011-04-01 | | | | | |
| TTBS | 2011-02-14 | 2011-03-16 | 2011-04-13 | | | | | |
| BSJ | 2011-03-16 | 2011-04-27 | 2011-06-09 | | | | | |
| CENAM (Pilot) | 2011-04-28 | | 2011-05-25 | | | | | |

Tabla 2. SIM.L-S6:2007 dates of reception and shipment of artifacts and reception of results by the pilot laboratory.

4. Comparison Artifacts

A total of 14 grade K (according to [1]) rectangular GB were selected for the exercise. Seven steel GB and seven ceramics GB covering the range of short GB (from 0.5 mm to 100 mm). The specifications on the GB are shown in **tables 3** and **4**. The associated Coefficients of Thermal Expansion (CET) shown in the tables are those quoted by the manufacturer.

| Nominal Length (mm) | Serial Number | Coefficient of Thermal Expansion (10 ⁻⁶ K ⁻¹) | Manufacturer |
|---------------------|---------------|---|--------------|
| 1.000 5 | 010223 | 10.9 ± 1 | Mitutoyo |
| 5 | 000482 | 10.9 ± 1 | Mitutoyo |
| 7 | 010764 | 10.9 ± 1 | Mitutoyo |
| 10 | 001329 | 10.9 ± 1 | Mitutoyo |
| 50 | 012254 | 10.9 ± 1 | Mitutoyo |
| 75 | 010630 | 10.9 ± 1 | Mitutoyo |
| 100 | 010850 | 10.9 ± 1 | Mitutoyo |

Table 3. Steel Gauge Blocks.

| Nominal Length (mm) | Serial Number | Coefficient of Thermal Expansion (10 ⁻⁶ K ⁻¹) | Manufacturer |
|---------------------|---------------|---|--------------|
| 1.000 5 | 000288 | 9.3 ± 1 | Mitutoyo |
| 5 | 051836 | 9.3 ± 1 | Mitutoyo |
| 7 | 010323 | 9.3 ± 1 | Mitutoyo |
| 10 | 052351 | 9.3 ± 1 | Mitutoyo |
| 50 | 011002 | 9.3 ± 1 | Mitutoyo |
| 75 | 010370 | 9.3 ± 1 | Mitutoyo |
| 100 | 010773 | 9.3 ± 1 | Mitutoyo |

Table 4. Ceramics Gauge Blocks.

5. Measurement Protocol

Detailed instructions were included in the technical protocol. Participants were invited to perform the measurements according to their own calibration procedures, used to calibrate the GB of their customers.

The measurement was performed in all cases with a double probe GB comparator and in the vertical position as indicated in [1]. The method determines the difference in central length, l_c , of two GB of same nominal length set beside in the comparator platen as illustrated in **Figure 1**. The first GB is the laboratory's reference GB, calibrated by optical interferometry; and the test GB which is under circulation.




Figure 1. Illustration of the calibration method of GB by mechanical comparison showing the different variables of influence (taken from [9]).

6. Measuring Instruments

All participants measured with double probe electromechanical GB comparators with a resolution of 10 nm. Table 5 shows the makes, models and characteristics of the different instruments.

| NMI | Manufacturer | Model | Measuring range mm | Traceability | Temperature variation range during measurements (°C) |
|----------|--------------|---------------|--------------------------|--|--|
| CENAM | TESA | TESA- UPC | 0 - 102 | To SI standards of CENAM via GB calibrated by interferometry | 19.32 – 19.60 |
| NPLI | Mahr | Not indicated | 0 – 175 | Not indicated | 19.5 – 20.5 |
| СМІ | TESA | TESA- UPC | 0 – 100 | To the Czech National Standard of Length (He-Ne/I2 633nm, He-Ne/I2 543.5nm, fs comb) | Not indicated |
| NIST | Mahr/Federal | 130B-24 | 0 -102 | NIST maintained lodine- Stabilized Laser | 20.08 – 20.17 |
| INMETRO | TESA | TESA- UPC | 0 - 102 | To SI standards of INMETRO via GB calibrated by interferometry | 20.0 – 20.5 |
| INTI | Mahr | 826E | 0 - 175 | To SI standards of INTI via GB calibrated by interferometry | 20 ± 0.5 |
| DICTUC | TESA | TESA- UPC | 0 - 100 | To SI standards of PTB via GB calibrated by interferometry | Steel 19.91 – 20.05 Ceramics 19.85 – 20.16 |
| LATU | Mahr | 826 | 0 - 100 | Comparator to SI standards of PTB via GB calibrated by interferometry and to SI standards of CENAM via GB calibrated by interferometry | 19.68 – 20.50 |
| INDECOPI | Mahr | 826 PC | 0 – 175 | To SI standards of CENAM via GB calibrated by interferometry | 20.0 ± 0.5 |
| IBMETRO | Steinmayer | EMP II | 0 – 100 | To SI standards of PTB via GB grade 0 | 19.95 – 20.15 |
| INEN | Mahr | 826 | 0 - 170 | Not indicated, only mentioned that their GB are calibrated by interferometry | 19 – 21 |
| SIC | TESA | UPD | 0 - 500 | To SI standards of METAS via GB calibrated by interferometry | 20 – 20.3 |
| CENAMEP | Mitutoyo | GBCD- 250 | 0 - 250 | To SI standards of CENAM via GB calibrated by interferometry | 20.3 – 20.6 |
| LACOMET | TESA | TESA- UPC | 0.5 - 100 | To SI standards of PTB and CENAM via GB calibrated by interferometry | 0.34 |
| TTBS | TESA | TESA- UPC | 0.5 - 100 | To SI standards of METAS and NPL via GB calibrated by interferometry | 0.018 |
| JBS | Mahr-Federal | 2247386 | 1 - 100 | To SI standards of NIST | 19.60 – 20.20 |

Table 5. GB comparators, measurement range, traceability and temperature variation of the participant laboratories.


7. State and Behavior of Artifacts

7. 1 State of the Artifacts upon Reception

The participants were to inspect the state of the artifacts upon reception and inform the pilot according to the protocol. Although the selected GB were not brand new, they were in good conditions. They suffered some damage after the circulation, but the stability and the results obtained in the comparison prove the damages did not hamper or alter the measurements and the pilot laboratory was able to wring them all to a measurement platen without problems after circulation. **Figures 1** through **11** show the physical conditions of some of the steel GB upon reception at the pilot laboratory at the end of the exercise. The steel GB ended-up with quite a few scratches and specifically, the 5 mm GB, also presented rust spots.



Figures 1, 2 and **3.** Aspect of the measuring faces of the 1.000 5, 5 and 7 mm Steel GB at the end of circulation.




Figures 4, 5 and **6.** Aspect of the measuring faces of the 10, 50 and 75 mm Steel GB at the end of circulation.



Figure 7. Aspect of the measuring faces of the 100 mm Steel GB at the end of circulation.

Four out of the seven ceramic GB suffered some damage on one of their measuring faces as shown in **Figures 8** to **11**, which consisted of burs or chipped edges. However, this condition did not hamper the wringing or caused any variation in length as it was proved by the control measurements performed at the end by the pilot laboratory by interferometry as well as by mechanical comparison.




Figures 8, 9, 10 and **11.** Aspect of the measuring faces of the 10, 50, 75 and 100 mm Ceramic GB at the end of circulation.

7. 2 Stability of the Standards

The GB were measured by interferometry several times by the pilot laboratory to verify their stability: when they were purchased (2002), two years before starting the comparison (Nov. 2005), before circulating them (Nov. 2007), after the round by interferometry (April 2010) and at the end of the circulation (May 2011). **Table 6** shows the deviations from nominal length determined at these different occasions for the steel GB, including the stated values on the certificates of the manufacturer. **Graphs 1** through **7** show these values for each GB.

| | Nominal | Deviation from nominal value (nm) | | | | | | | | |
|------------------|----------------|-----------------------------------|------|------|------|------|------|--|--|--|
| Serial Number | length (mm) | Manufacturer certificate 2001 | 2002 | 2005 | 2007 | 2010 | 2011 | | | |
| 010223 | 1.000 5 | 0 | 5 | 3 | -4 | -9 | -2 | | | |
| 000482 | 5 | 40 | 14 | 11 | 35 | 20 | 27 | | | |
| 010764 | 7 | 30 | 19 | 13 | -5 | 1 | -10 | | | |
| 001329 | 10 | 50 | 31 | 22 | 37 | 21 | 31 | | | |
| 012254 | 50 | 60 | 46 | 3 | 7 | -3 | 8 | | | |
| 010630 | 75 | -50 | -54 | -104 | -100 | -107 | -106 | | | |
| 010850 | 100 | 20 | 18 | -50 | -51 | -64 | -50 | | | |

Table 6. Pilot Laboratory measured values of the steel GB at different occasions.









Table 7 shows the deviations from nominal length determined at these different occasions for the ceramics GB, including the stated values on the certificates of the manufacturer. **Graphs 8** through **14** show these values for each GB.


| | Nominal | Deviation from nominal value (nm) | | | | | | | | |
|------------------|----------------|-----------------------------------|------|-------------------------------|------|------|------|--|--|--|
| Serial Number | length (mm) | Manufacturer certificate 2001 | 2002 | Manufacturer certificate 2005 | 2007 | 2010 | 2011 | | | |
| 000288 | 1.0005 | 0 | -6 | | -14 | 7 | -17 | | | |
| 051836 | 5 | | | 13 | 12 | 8 | 14 | | | |
| 010323 | 7 | 50 | 46 | | 57 | 48 | 35 | | | |
| 052351 | 10 | | | 3 | -13 | -19 | -29 | | | |
| 011002 | 50 | 90 | 95 | | 139 | 117 | 97 | | | |
| 010370 | 75 | 100 | 110 | | 118 | 124 | 123 | | | |
| 010773 | 100 | -60 | -42 | | -34 | -36 | -41 | | | |


Table 7. Pilot Laboratory measured values of the ceramics *GB* in different occasions.

















8. Measurement Results of Participants

All laboratories sent their results by e-mail. All information was received on the specified formats from appendices A, B, C, D and E of the Technical Protocol.

8.1 Measurement of the Central Length

Tables 8 and **9** and **graphs 16** through **22**, show the deviations of the central length with respect to nominal values and the claimed standard measurement uncertainties of each participant for the seven steel GB. Additionally, **graph 15** shows the claimed standard uncertainties of all participants.

| Nominal | | Deviation (e_{ij}) from nominal length for Steel GB | | | | | | | | |
|---------|------|---|------|---------|------|------|--------|----------|--|--|
| Value | | | T | nm | | | ı | | | |
| mm | NPLI | CMI | NIST | INMETRO | INTI | LATU | DICTUC | INDECOPI | | |
| 1.000 5 | -20 | 20 | -27 | 20 | -12 | -35 | -15 | 20 | | |
| 5 | 10 | 60 | 27 | 62 | -4 | 4 | 5 | 20 | | |
| 7 | -40 | 20 | 2 | 21 | -38 | -12 | -35 | 10 | | |
| 10 | 10 | 70 | 39 | 24 | 12 | 39 | 20 | 40 | | |
| 50 | -30 | 80 | 15 | 11 | -11 | 13 | -10 | 0 | | |
| 75 | 60 | -50 | -107 | -93 | -127 | -126 | -110 | -80 | | |
| 100 | 10 | 40 | -67 | -60 | -66 | -68 | -60 | 10 | | |

Table 8A. Measurement results of the participants for the Steel GB.

| Nominal | | Deviation (e_{ij}) from nominal length for Steel GB | | | | | | | |
|---------|---------|---|------|---------|---------|------|--------------|-------|--|
| Value | | | | n | m | | | | |
| mm | IBMETRO | INEN | SIC | CENAMEP | LACOMET | TTBS | BSJ | CENAM | |
| 1.000 5 | -40 | -10 | -27 | -13 | 22 | 30 | Not reported | -2 | |
| 5 | 0 | 10 | -6 | -3 | 46 | 70 | 11 | 23 | |
| 7 | -40 | -30 | -38 | -13 | 27 | 20 | -37 | 6 | |
| 10 | 20 | 20 | -7 | 31 | 66 | 20 | 15 | 24 | |
| 50 | -10 | 10 | 0 | -10 | 36 | 200 | -26 | 14 | |
| 75 | -130 | -140 | -127 | -143 | -83 | -40 | -295 | -109 | |
| 100 | -70 | 40 | -23 | -112 | -68 | 290 | -47 | -33 | |


Table 8B. Measurement results of the participants for the Steel GB.


| Nominal Value | | Claimed standard uncertainties, $u(e_{ij})$, Steel GB nm | | | | | | | | |
|------------------|------|---|------|---------|------|------|--------|----------|--|--|
| mm | NPLI | CMI | NIST | INMETRO | INTI | LATU | DICTUC | INDECOPI | | |
| 1.000 5 | 26 | 23 | 13 | 29.5 | 23.0 | 26.0 | 27 | 27 | | |
| 5 | 28 | 23.1 | 13 | 29.8 | 23.0 | 26.0 | 27 | 27 | | |
| 7 | 29 | 23.1 | 13 | 29.8 | 23.0 | 26.0 | 27 | 27 | | |
| 10 | 30 | 23.3 | 13 | 29.8 | 23.0 | 26.0 | 27 | 27 | | |
| 50 | 53 | 29.2 | 15 | 35.2 | 35.0 | 38.0 | 29 | 36 | | |
| 75 | 67 | 35.5 | 20 | 41.4 | 46.0 | 40.0 | 34 | 45 | | |
| 100 | 81 | 42.7 | 30 | 49.1 | 58.0 | 47.0 | 36 | 55 | | |


Table 9A. Claimed standard uncertainties of the participants for the Steel GB.


| Nominal Value | | Claimed standard uncertainties, $u(e_{ij})$, Steel GB nm | | | | | | | |
|------------------|---------|---|-----|---------|---------|------|--------------|-------|--|
| mm | IBMETRO | INEN | SIC | CENAMEP | LACOMET | TTBS | BSJ | CENAM | |
| 1.000 5 | 34 | 55.4 | 23 | 12 | 25.3 | 52 | Not reported | 14.5 | |
| 5 | 35 | 57.2 | 23 | 12 | 25.4 | 53 | 28 | 14.7 | |
| 7 | 35 | 58 | 23 | 12 | 36.7 | 53 | 29 | 14.9 | |
| 10 | 36 | 59.3 | 24 | 13 | 36.9 | 77 | 31 | 15.2 | |
| 50 | 44 | 76.5 | 34 | 35 | 43.7 | 160 | 71 | 27.6 | |
| 75 | 50 | 87.3 | 44 | 51 | 43.8 | 170 | 100 | 38.1 | |
| 100 | 55 | 98 | 55 | 66 | 54.4 | 170 | 130 | 49.2 | |


Table 9B. Claimed standard uncertainties of the participants for the Steel GB.




















Tables 10 and **11** and **graphs 24** through **30**, show the deviations of the central length with respect to nominal values and their claimed standard measurement uncertainties of each participant for the seven ceramic GB. **Graph 23** shows the claimed standard uncertainties of the participants.

| Nominal | | Deviation (e_{ij}) from nominal length for Ceramics GB | | | | | | | | | |
|---------|------|--|------|---------|------|------|--------|----------|--|--|--|
| Value | | nm | | | | | | | | | |
| mm | NPLI | CMI | NIST | INMETRO | INTI | LATU | DICTUC | INDECOPI | | | |
| 1.000 5 | -40 | 0 | -27 | -32 | -9 | -37 | -17 | -15 | | | |
| 5 | 0 | 30 | 14 | -14 | 10 | -29 | -17 | -15 | | | |
| 7 | 10 | 50 | 58 | 34 | 35 | 13 | 13 | 25 | | | |
| 10 | -80 | 0 | -9 | -24 | -15 | -18 | -32 | -25 | | | |
| 50 | 20 | 140 | 116 | 97 | 89 | 103 | 65 | 55 | | | |
| 75 | 40 | 150 | 148 | 89 | 122 | 111 | 101 | 135 | | | |
| 100 | -80 | 10 | -5 | -55 | 16 | -53 | -65 | -15 | | | |

Table 10A. Measurement results of the participants for Ceramics GB.

| Nominal | | Devi | ation (e | _{ij}) from nom | inal length f | or Cerami | es <i>GB</i> | |
|---------|---------|------|----------|--------------------------|---------------|--------------|--------------|-------|
| Value | | | | | nm | | | |
| mm | IBMETRO | INEN | SIC | CENAMEP | LACOMET | TTBS | BSJ | CENAM |
| 1.000 5 | 310 | -10 | -5 | -12 | -9 | 60 | Not reported | -10 |
| 5 | 340 | -20 | -10 | -1 | -16 | 80 | -36 | -6 |
| 7 | 370 | 20 | 32 | 38 | 44 | 90 | -18 | 65 |
| 10 | 320 | -10 | -40 | -18 | -15 | Not reported | -68 | -20 |
| 50 | 460 | 140 | 138 | 103 | 84 | 480 | -16 | 100 |
| 75 | 480 | 170 | 155 | 164 | 102 | 480 | -140 | 127 |
| 100 | 360 | 40 | 40 | 2 | 8 | 660 | -195 | -12 |


Table 10B. Measurement results of the participants for Ceramics GB.


| Nominal Value | | Clai | med stan | | tainties, <i>u</i> nm | $u(e_{ij})$, Ceram | ics <i>GB</i> | | | | | |
|------------------|------|---|----------|------|--------------------------|---------------------|---------------|----|--|--|--|--|
| mm | NPLI | NPLI CMI NIST INMETRO INTI LATU DICTUC INDECOPI | | | | | | | | | | |
| 1.000 5 | 26 | 23 | 13 | 32.5 | 23 | 26 | 27 | 27 | | | | |
| 5 | 28 | 23.1 | 13 | 33.2 | 23 | 26 | 27 | 27 | | | | |
| 7 | 29 | 23.1 | 13 | 32.7 | 23 | 26 | 27 | 27 | | | | |
| 10 | 30 | 23.3 | 13 | 32.7 | 23 | 26 | 27 | 27 | | | | |
| 50 | 53 | 29.2 | 15 | 38.1 | 33 | 36 | 28 | 38 | | | | |
| 75 | 67 | 35.5 | 20 | 44 | 42 | 45 | 35 | 48 | | | | |
| 100 | 81 | 42.7 | 30 | 50.6 | 52 | 47 | 34 | 59 | | | | |


Table 11A. Claimed standard uncertainties of the participants for Ceramics GB.


| Nominal | | Clair | med sta | ndard uncer | tainties, <i>u</i> (<i>e</i> ; | _j), Ceramio | es GB | | | |
|---------|---------|---|---------|-------------|---------------------------------|-------------------------|--------------|------|--|--|
| Value | | | | | nm | | | | | |
| mm | IBMETRO | BMETRO INEN SIC CENAMEP LACOMET TTBS BSJ CE | | | | | | | | |
| 1.000 5 | 34 | 55.4 | 23 | 12 | 25.3 | 58 | Not reported | 19.0 | | |
| 5 | 35 | 57.2 | 23 | 12 | 25.4 | 58 | 39 | 19.2 | | |
| 7 | 35 | 58 | 23 | 12 | 25.5 | 58 | 40 | 19.4 | | |
| 10 | 36 | 59.3 | 24 | 13 | 25.7 | Not reported | 43 | 19.9 | | |
| 50 | 44 | 76.5 | 34 | 34 | 43.5 | 160 | 99 | 34.5 | | |
| 75 | 50 | 87.3 | 44 | 50 | 43 | 170 | 140 | 47.1 | | |
| 100 | 56 | 98 | 55 | 66 | 55.3 | 180 | 182 | 60.6 | | |


Table 11B. Claimed standard uncertainties of the participants for Ceramics GB.




















9. Reference Value (RV) and equations to determine the performance of participants

All usual parameters of the central tendency were calculated: the median, the simple mean and the inverse-variance weighted mean. All of these values appear in **Annex A**. However, the Supplementary Comparison Reference Values were obtained from the Regional Key Comparison SIM.L-K1:2007, as the artifacts were the same. These RV were calculated as the simple mean for all consistent results of the interferometric comparison [2].

The Reference Values, \bar{e}_j and their Expanded Uncertainties, $U(\bar{e}_j)$ for the different GB_j of both materials are shown in **Table 13**.

| Supplen | nentary Comp | arison Refe | rence Values | (RV) | | |
|--------------|--------------------------|----------------|--------------------------|----------------|--|--|
| Nominal | Ste | el | Ceramics | | | |
| Length mm | Ref. Value, $ar{e}_j$ | $U(\bar{e}_j)$ | Ref. Value, $ar{e}_j$ | $U(\bar{e}_j)$ | | |
| 1.005 | -9.3 | 8.2 | -3.9 | 7.8 | | |
| 5 | 26.4 | 8.3 | 10.6 | 7.9 | | |
| 7 | -2.6 | 8.3 | 51.7 | 7.9 | | |
| 10 | 36.2 | 7.8 | -14.3 | 8.1 | | |
| 50 | 4.9 | 12.5 | 105.9 | 11.3 | | |
| 75 | -105.6 | 17.0 | 136.2 | 13.9 | | |
| 100 | -42.0 | 17.1 | -23.1 | 15.6 | | |

Table 13. Reference values (simple mean of largest sub-set of consistent results of SIM.L-K1:2007 comparison) as deviations from Nominal Value and corresponding Expanded Uncertainty for both steel and ceramic *GB*. All values in nanometers.

For each laboratory, i, which measures each gauge block, j, let the measured deviation from nominal length be denoted by d_{ij} and calculated as,

$$d_{ij} = e_{ij} - \bar{e}_i \tag{1}$$

Statistical consistency of the results with their associated uncertainties can be verified by calculating the normalized error E_n .

$$E_n = \frac{|d_{ij}|}{\sqrt{U^2(e_{ij}) + U^2(\overline{e}_j)}}$$
 (2)

If E_n is greater than 1 it is considered that the result is inconsistent.

10. Participants Results of the Comparison

The reported measurement results were analyzed by simple statistical means to allow identification of any significant bias. **Tables 14** and **15** show the differences of the results of the participants with respect to the **RV** of each GB j, d_{ij} , and the corresponding Normalized Error, E_n calculated from equation (2). Note that the uncertainties in this equation are expanded uncertainties.

| NMI (i→) | NF | PLI | CI | МІ | NI | ST | INME | TRO | IN | TI | LA | TU |
|------------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|
| Nominal Length (j↓) | d_{ij} | \boldsymbol{E}_n |
| 1.000 5 | -10.7 | 0.20 | 29.3 | 0.63 | -17.7 | 0.65 | 29.3 | 0.49 | -2.7 | 0.06 | -25.7 | 0.49 |
| 5 | -16.4 | 0.29 | 33.6 | 0.72 | 0.6 | 0.02 | 35.6 | 0.59 | -30.4 | 0.65 | -22.4 | 0.43 |
| 7 | -37.4 | 0.64 | 22.6 | 0.48 | 4.6 | 0.17 | 23.6 | 0.39 | -35.4 | 0.76 | -9.4 | 0.18 |
| 10 | -26.2 | 0.43 | 33.8 | 0.72 | 2.8 | 0.10 | -12.2 | 0.20 | -24.2 | 0.52 | 2.8 | 0.05 |
| 50 | -34.9 | 0.33 | 75.1 | 1.26 | 10.1 | 0.31 | 6.1 | 0.09 | -15.9 | 0.22 | 8.1 | 0.11 |
| 75 | 165.6 | 1.23 | 55.6 | 0.76 | -1.4 | 0.03 | 12.6 | 0.15 | -21.4 | 0.23 | -20.4 | 0.25 |
| 100 | 52 | 0.32 | 82 | 0.94 | -25 | 0.40 | -18 | 0.18 | -24 | 0.20 | -26 | 0.27 |

Table 14A. Deviation from reference value for each GB, d_{ij} (nanometers) and Normalized Error, E_n , of the **Steel GB** for the first six participants.

| NMI (i→) | DIC | TUC | INDE | СОРІ | IBME | TRO | INI | EN | SI | iC | CENA | MEP |
|------------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|
| Nominal Length (j↓) | d_{ij} | \boldsymbol{E}_n |
| 1.000 5 | -5.7 | 0.10 | 29.3 | 0.54 | -30.7 | 0.45 | -0.7 | 0.01 | -17.7 | 0.38 | -3.7 | 0.15 |
| 5 | -21.4 | 0.39 | -6.4 | 0.12 | -26.4 | 0.37 | -16.4 | 0.14 | -32.4 | 0.69 | -29.4 | 1.16 |
| 7 | -32.4 | 0.59 | 12.6 | 0.23 | -37.4 | 0.53 | -27.4 | 0.24 | -35.4 | 0.76 | -10.4 | 0.41 |
| 10 | -16.2 | 0.30 | 3.8 | 0.07 | -16.2 | 0.22 | -16.2 | 0.14 | -43.2 | 0.89 | -5.2 | 0.19 |
| 50 | -14.9 | 0.25 | -4.9 | 0.07 | -14.9 | 0.17 | 5.1 | 0.03 | -4.9 | 0.07 | -14.9 | 0.21 |
| 75 | -4.4 | 0.06 | 25.6 | 0.28 | -24.4 | 0.24 | -34.4 | 0.20 | -21.4 | 0.24 | -37.4 | 0.36 |
| 100 | -18 | 0.24 | 52 | 0.47 | -28 | 0.25 | 82 | 0.42 | 19 | 0.17 | -70 | 0.53 |

Table 14B. Deviation from reference value for each GB, d_{ij} (nanometers) and Normalized Error, E_n of the **Steel GB** for the next six participants.

| NMI (i→) | LACOMET | | TTBS | | В | SJ | CENAM | | |
|------------------------|----------|--------------------|----------|--------------------|--------------|--------------------|----------|--------------------|--|
| Nominal Length (j↓) | d_{ij} | \boldsymbol{E}_n | d_{ij} | \boldsymbol{E}_n | d_{ij} | \boldsymbol{E}_n | d_{ij} | \boldsymbol{E}_n | |
| 1.000 5 | 31.3 | 0.61 | 39.3 | 0.38 | Not reported | Not reported | 7.3 | 0.24 | |
| 5 | 19.6 | 0.38 | 43.6 | 0.41 | -15.4 | 0.27 | -3.4 | 0.11 | |
| 7 | 29.6 | 0.40 | 22.6 | 0.21 | -34.4 | 0.59 | 8.6 | 0.28 | |
| 10 | 29.8 | 0.40 | -16.2 | 0.11 | -21.2 | 0.34 | -12.2 | 0.39 | |
| 50 | 31.1 | 0.35 | 195.1 | 0.61 | -30.9 | 0.22 | 9.1 | 0.16 | |
| 75 | 22.6 | 0.25 | 65.6 | 0.19 | -189.4 | 0.94 | -3.4 | 0.04 | |
| 100 | -26 | 0.24 | 332 | 0.98 | -5 | 0.02 | 9 | 0.09 | |

Table 14C. Deviation from reference value for each GB, d_{ij} (nanometers) and Normalized Error, E_{nv} of the **Steel GB** for the last four participants.

| NMI (i→) | NF | PLI | C | МІ | NI | ST | INME | TRO | IN | TI | LA | TU |
|------------------------|----------|-------|----------|-------|----------|-------------|----------|-------|----------|-------|----------|-------|
| Nominal Length (j↓) | d_{ij} | E_n | d_{ij} | E_n | d_{ij} | E_n E_n | d_{ij} | E_n | d_{ij} | E_n | d_{ij} | E_n |
| 1.000 5 | -36.1 | 0.69 | 3.9 | 0.08 | -23.1 | 0.85 | -28.1 | 0.43 | -5.1 | 0.11 | -33.1 | 0.63 |
| 5 | -10.6 | 0.19 | 19.4 | 0.41 | 3.4 | 0.13 | -24.6 | 0.37 | -0.6 | 0.01 | -39.6 | 0.75 |
| 7 | -41.7 | 0.71 | -1.7 | 0.04 | 6.3 | 0.23 | -17.7 | 0.27 | -16.7 | 0.36 | -38.7 | 0.74 |
| 10 | -65.7 | 1.09 | 14.3 | 0.30 | 5.3 | 0.19 | -9.7 | 0.15 | -0.7 | 0.01 | -3.7 | 0.07 |
| 50 | -85.9 | 0.81 | 34.1 | 0.57 | 10.1 | 0.32 | -8.9 | 0.12 | -16.9 | 0.25 | -2.9 | 0.04 |
| 75 | -96.2 | 0.71 | 13.8 | 0.19 | 11.8 | 0.28 | -47.2 | 0.53 | -14.2 | 0.17 | -25.2 | 0.28 |
| 100 | -56.9 | 0.35 | 33.1 | 0.38 | 18.1 | 0.29 | -31.9 | 0.31 | 39.1 | 0.37 | -29.9 | 0.31 |

Table 15A. Deviation from reference value for each GB, d_{ij} (nanometers) and Normalized Error, E_{n} of the **Ceramics GB** for the first six participants.

| NMI (i→) | DIC | TUC | INDE | СОРІ | IBME | TRO | IN | EN | S | IC | CENA | AMEP |
|------------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|----------|--------------------|
| Nominal Length (j↓) | d_{ij} | \boldsymbol{E}_n |
| 1.000 5 | -13.1 | 0.24 | -11.1 | 0.20 | 313.9 | 4.59 | -6.1 | 0.05 | -1.1 | 0.02 | -8.1 | 0.32 |
| 5 | -27.6 | 0.51 | -25.6 | 0.47 | 329.4 | 4.68 | -30.6 | 0.27 | -20.6 | 0.44 | -11.6 | 0.46 |
| 7 | -38.7 | 0.71 | -26.7 | 0.49 | 318.3 | 4.52 | -31.7 | 0.27 | -19.7 | 0.42 | -13.7 | 0.54 |
| 10 | -17.7 | 0.32 | -10.7 | 0.20 | 334.3 | 4.61 | 4.3 | 0.04 | -25.7 | 0.53 | -3.7 | 0.14 |
| 50 | -40.9 | 0.72 | -50.9 | 0.66 | 354.1 | 3.99 | 34.1 | 0.22 | 32.1 | 0.47 | -2.9 | 0.04 |
| 75 | -35.2 | 0.49 | -1.2 | 0.01 | 343.8 | 3.41 | 33.8 | 0.19 | 18.8 | 0.21 | 27.8 | 0.28 |
| 100 | -41.9 | 0.60 | 8.1 | 0.07 | 383.1 | 3.39 | 63.1 | 0.32 | 63.1 | 0.57 | 25.1 | 0.19 |

Table 15B. Deviation from reference value for each GB, d_{ij} (nanometers) and Normalized Error, E_n of the **Ceramics GB** for the next six participants.

| NMI (i→) | LACOMET | | TTBS | | BS | J | CENAM | | |
|------------------------|----------|--------------------|--------------|--------------------|--------------|--------------------|----------|--------------------|--|
| Nominal Length (j↓) | d_{ij} | \boldsymbol{E}_n | d_{ij} | \boldsymbol{E}_n | d_{ij} | \boldsymbol{E}_n | d_{ij} | \boldsymbol{E}_n | |
| 1.000 5 | -5.1 | 0.10 | 63.9 | 0.55 | Not reported | Not reported | -6.1 | 0.16 | |
| 5 | -26.6 | 0.52 | 69.4 | 0.60 | -46.6 | 0.59 | -16.6 | 0.42 | |
| 7 | -7.7 | 0.15 | 38.3 | 0.33 | -69.7 | 0.87 | 13.3 | 0.34 | |
| 10 | -0.7 | 0.01 | Not reported | Not reported | -53.7 | 0.62 | -5.7 | 0.14 | |
| 50 | -21.9 | 0.25 | 374.1 | 1.17 | -121.9 | 0.61 | -5.9 | 0.08 | |
| 75 | -34.2 | 0.39 | 343.8 | 1.01 | -276.2 | 0.99 | -9.2 | 0.10 | |
| 100 | 31.1 | 0.28 | 683.1 | 1.90 | -171.9 | 0.47 | 11.1 | 0.09 | |

Table 15C. Deviation from reference value for each GB, d_{ij} (nanometers) and Normalized Error, E_{n} of the **Ceramics GB** for the last four participants.

Table 16 shows the Root Mean Square (RMS) values of the deviations of the participants, δ_{RMS} , with respect to the RV. It gives a general idea of the deviations of each participant with respect to RV. It is determined as:

$$\delta_{RMS} = \sqrt{\frac{\sum_{i=1}^{n} (e_{ij} - \overline{e}_{j})^{2}}{n}}$$
(3)

Where:

 e_{ii} – Deviation from nominal of laboratory i on GB j,

 \bar{e}_i – RV of GB j,

n – Number of GB

| NMI | $\delta_{	extit{RMS}}$ | (nm) |
|----------|------------------------|-------------|
| INIVII | Steel GB | Ceramics GB |
| NPLI | 69.5 | 62.5 |
| CMI | 52.2 | 20.9 |
| NIST | 12.4 | 13.0 |
| INMETRO | 21.9 | 27.1 |
| INTI | 24.1 | 18.2 |
| LATU | 18.6 | 28.6 |
| DICTUC | 18.4 | 32.5 |
| INDECOPI | 25.2 | 24.7 |
| IBMETRO | 26.5 | 340.3 |
| INEN | 36.3 | 34.4 |
| SIC | 27.6 | 31.3 |
| CENAMEP | 32.8 | 16.1 |
| LACOMET | 27.5 | 22.1 |
| TTBS | 149.7 | 350.0 |
| BSJ | 80.8 | 147.5 |
| CENAM | 8.1 | 10.5 |

Table 16. RMS Values of the deviations with respect to the RV, δ_{RMS} , of the participants.

In this case, as the RV is determined external to the comparison, it is not possible calculate Birge ratio.

11. Discussion and Conclusions

11.1 Discussion

- The comparison was linked to a previous interferometric comparison that measured the same artifacts. This was an advantage as the RV were obtained from the interferometric stage providing low uncertainty RV for the Mechanical Compariosn exercise. Not only were the RV obtained by a metrological superior technique, but it was the result of the measurement of several participants that measured by this technique.
- In the second stage of circulation which included only those NMI that measured exclusively by mechanical comparison (10 laboratories), the timing of circulation (from 2010-04-15 to 2011-04-28) was short thanks to the hand-delivery of the artifacts to the following participant. We would like to thank the participants for having taken this trouble and we suggest adopting this transport option whenever possible as it reduces the time of circulation and the risk of damage to the artifacts during transport.

- There was also some time saving during circulation in the first stage, as the NMI that
 participated in both exercises received the artifacts only once and measured by both
 techniques during the same period.
- Declared standard uncertainties among participants spread over a 6 fold range, going from 10 nm to 60 nm for the shortest GB and from 25 nm to 180 nm for the 100 mm GB.
- A few participants had the same traceability source because their master GB were calibrated at a same laboratory. However, we consider the influence of these correlations minimal and they were not taken in account in the present analysis.

11.2 Conclusions

- From Section 7 we observe that there were no appreciable changes on the
 measurements performed by the pilot laboratory of the ensemble of the GB of both
 materials over the last five years. Even though some drift may be appreciated on the
 steel GB during their first years of their history, the values shown prove they reached
 stability since 2005 approximately. Therefore, it can be assumed that the artifacts
 behaved adequately during the comparison exercise and that the exercise was valid.
- It may not be asserted that there was more consistency in one material or another and results were similar for both materials.
- From the comparison of the simple mean and weighted mean presented in Appendix A with the RV for all GB, we observe that both means are always pretty close to the RV. We do not identify either any systematic effect between the Interferometric mean or RV and the mechanical comparison mean.
- Once Expanded Uncertainties are considered, the performance of most of the participants for the steel GB was good, even though three results had normalized errors greater than one: NPLI on the 75 mm, CMI on the 50 mm and CENAMEP on the 5 mm; and BSJ did not present results for the 1.000 5 mm GB.
- IBMETRO presented inconsistent results for the ceramic GB. As they obtained
 consistent results for the steel GB, we presumed their "raw" measurements of the
 Ceramic GB were probably good, but that they applied a wrong correction for
 comparing GB of two different materials. In effect, they informed us after circulation of
 DRAFT A, they made such a mistake and they are already amending the
 miscalculation in the procedures of their Quality Management System.
- Also for the ceramic GB TTBS obtained inconsistent results for the three longer ones;
 and TTBS and BSJ did not present results for two nominal values.
- For the rest of the participants their results are judged satisfactory which proves their technical competence.

12. Acknowledgements

We would like to acknowledge:

 SIM WG.4 Length and SIM Technical Committee for having funded the purchase of the fourteen GB to carry this comparison.

 In an anonymous way, the technicians and colleagues from our different institutions that contributed directly or indirectly to the measurements of the artifacts in this comparison.

13. References

- **1. ISO 3650:1998(E)**, Geometrical Product Specification (GPS) Length Standards Gauge Blocks, International Organization for Standardization, Geneva, Switzerland.
- **2. C. Colín, M. Viliesid, et.al**. SIM Regional Key Comparison SIM.L-K1:2007. Calibration of Gauge Blocks by Optical Interferometry. June 2012.
- **3. Thalmann R.**, *CCL-K1 Final report, Calibration of gauge blocks by interferometry*, Wabern, Switzerland, November 2000. At *BIPM* website, http://www.bipm.fr.
- **4.** Viliesid M., Comparison CCL-K6 "Calibration of Coordinate Measuring Machine (CMM) Two-dimensional (2-D) Artifacts (Ball plates and Bore Plates)" Final Report, 2008- 10-27 At BIPM website, http://www.bipm.fr.
- **5.** Cox, M.G., 2002, The evaluation of key comparison data, Metrologia, 2003, **39**, pp 589-595.
- **6. Beissner, K.**, 2002, On a measure of consistency in comparison measurements, Metrologia, 2003, **39**, pp 59-63.
- Kacker, R., Datla, R., Parr, A., 2002, Combined result and associated uncertainty from interlaboratory evaluations based on the ISO Guide, Metrologia, 2003, 39, pp 279-293.
- 8. Decker J., Ulrich A., Lapointe A., Viliesid M., Pekelsky J. R., Two-part Study towards Lowest Uncertainty Calibration O Ceramic Gauge Blocks: Interferometry And Mechanical Comparison Techniques, in Recent Developments in Traceable Dimensional Measurements, Proceedings of SPIE The International Society for Optical Engineering, Vol. 4401 presented at Munich, Germany, June 2001.
- 9. Decker J., Pekelsky J. R., Uncertainty of Gauge Block Calibration by Mechanical Comparison: A Worked Example Case 1: Gauges of Like Materials, NRC Doc 39998, presented at National Conference of Standards Laboratories, Ottawa, Ontario, 1996.

Annex A Calculation of Alternate Statistical Parameters.

| | Steel gauş | ge blocks / | Nominal le | ength, mm | | | |
|----------------------------------|----------------------|-------------|------------|-----------|-----------------|-----------------|------------------|
| Statistical estimator | 1.000 5 ¹ | 5 | 7 | 10 | 50 ² | 75 ³ | 100 ⁴ |
| Reference Value, RV | -9.3 | 26.4 | -12.5 | 36.2 | 4.9 | -105.6 | -42.0 |
| Standard uncertainty | 4.1 | 4.2 | 4.2 | 3.9 | 6.3 | 8.5 | 8.6 |
| Simple arithmetic mean | -5.9 | 20.9 | -11.1 | 27.7 | 0.1 | -104.6 | -50.3 |
| Standard uncertainty | 7.7 | 7.6 | 7.8 | 8.7 | 11.7 | 17.1 | 17.6 |
| Birge Ratio | 0.77 | 0.83 | 0.83 | 0.57 | 0.40 | 0.50 | 0.54 |
| Weighted mean | -10.2 | 16.9 | -8.8 | 29.7 | 5.4 | -104.9 | -54.9 |
| Standard uncertainty | 5.5 | 5.4 | 5.5 | 5.7 | 8.6 | 10.5 | 13.8 |
| Birge Ratio | 0.87 | 0.97 | 0.94 | 0.76 | 0.44 | 0.62 | 0.56 |
| Median | -11.0 | 10.5 | -12.5 | 22.0 | 0.0 | -109.5 | -60.0 |
| Observed chi-squared | 10.7 | 17.3 | 14.6 | 10.1 | 2.6 | 5.0 | 4.6 |
| Degrees of freedom | 14 | 15 | 15 | 15 | 13 | 13 | 12 |
| $Pr\{\chi^2(v) > \chi^2_{obs}\}$ | 0.71 | 0.30 | 0.94 | 0.82 | 1.00 | 0.976 | 0.97 |
| Reduced chi-squared | 0.76 | 1.16 | 0.97 | 0.67 | 0.20 | 0.38 | 0.39 |

 $\ensuremath{^{1}}$ BSJ not considered in the statistical parameter calculations as they did not measure.

². CMI and TTBS were eliminated from the statistical parameter calculations.
³. NPLI and BSJ were eliminated from the statistical parameter calculations.

⁴ CMI, INEN and TTBS were eliminated from the statistical parameter calculations.

| Ceramics gauge blocks / Nominal length, mm | | | | | | | |
|--|----------------------|----------------|-----------------------|-----------------|-----------------|-------|------------------|
| Statistical estimator | 1.000 5 ⁵ | 5 ⁶ | 7 ⁶ | 10 ⁷ | 50 ⁸ | 75° | 100 ⁸ |
| Reference Value, RV | -3.9 | 10.6 | 51.7 | -14.3 | 105.9 | 136.2 | -23.1 |
| Standard uncertainty | 3.9 | 4.0 | 4.0 | 4.1 | 5.7 | 7.0 | 7.8 |
| Simple arithmetic mean | -11.6 | -2.0 | 33.9 | -26.7 | 96.2 | 131.2 | -13.0 |
| Standard uncertainty | 8.2 | 8.2 | 8.2 | 8.0 | 11.2 | 13.7 | 16.3 |
| Birge Ratio | 0.78 | 0.89 | 0.82 | 0.75 | 0.87 | 0.56 | 0.66 |
| Weighted mean | -16.1 | -0.5 | 39.8 | -20.6 | 102.8 | 132.4 | -17.5 |
| Standard uncertainty | 5.8 | 5.7 | 5.7 | 5.9 | 8.6 | 11.1 | 13.6 |
| Birge Ratio | 0.64 | 0.79 | 0.86 | 0.77 | 0.91 | 0.62 | 0.73 |
| Median | -11.0 | -10.0 | 34.0 | -19.0 | 100.0 | 131.0 | -5.0 |
| Observed chi-squared | 9.8 | 12.5 | 14.7 | 8.9 | 10.1 | 4.4 | 6.6 |
| Degrees of freedom | 13 | 14 | 14 | 13 | 12 | 11 | 12 |
| $Pr\{\chi^2(v) > \chi^2_{obs}\}$ | 0.71 | 0.57 | 0.40 | 0.78 | 0.61 | 0.96 | 0.88 |
| Reduced chi-squared | 0.75 | 0.89 | 1.05 | 0.69 | 0.84 | 0.40 | 0.55 |

 $^{^{\}rm 5}$ IBMETRO and BSJ were not considered for the statistical parameter calculations.

⁶ IBMETRO was not considered for the statistical parameter calculations.

⁷ IBMETRO and TTBS not considered for the statistical parameter calculations.

⁸ IBMETRO, TTBS and BSJ not considered for the statistical parameter calculations.

⁹ NPLI, IBMETRO, TTBS and BSJ were eliminated from the statistical parameter calculations.